

Elcometer Limited

Edge Lane, Manchester
M43 6BU. England
Tel: +44 (0)161 371 6000
Fax: +44 (0)161 371 6010
sales@elcometer.com
www.elcometer.com
REGISTERED NO. 1729726 ENGLAND
Accredited to ISO 9001
ISO 14001

Declaration of Conformity

The Directives covered by this declaration:

2014/30/EU Electromagnetic Compatibility

Declares that the product(s): Elcometer NDT UG20DL Underwater Thickness Gauge

Part Number(s): UG20DL

Product Option(s): TL-24030-4, TL-24031, TL-24032, TX2M25EL-2, TX3M50EP-3,

TX5M00EL-2, TX5M00EP-8, TX5M00EP-9

This declaration of conformity is issued under the sole responsibility of Elcometer Limited. The product identified above complies with the requirements of the above EU Directive by meeting the following standards:

EN 61326-1:2013 ¹	Electrical equipment for measurement, control
IEC 61326-1:2012	and laboratory use – EMC requirements
Class A ² , Group 1 ³ ISM	Part 1 General requirements.

¹ CKC Laboratories, INC. Test Report 90334-5

((

Signed:
M. C. Sellars

Manchester, UK Date: 4th July 2016

Authority: Managing Director

² Class A product: Suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

³ Group 1 ISM product: Product in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.